

Acoustic Calibrator - Class 2, 114 dB



Features

- IEC 60942:2003 Class 2
- ANSI S1.40:2006
- Sound pressure level 114 dB
- Frequency 1 kHz
- Fits ½" microphones

Applications

- Calibration of noise dosimeters
- Calibration check of occupational noise meter
- Field calibration checks
- Required by many noise regulations

Overview

The CR518 Acoustic Calibrator meets ANSI S1.40:2006 and IEC 60942:2003 to Class 2. It is intended for use with noise dosimeters and occupational noise meters that may require the higher calibration level.

This calibrator is included with the doseBadge Professional range of noise dosimeter kits, along with the doseBadges, charger, carrying case and software.

Using It

As soon as the calibrator is switched on it emits a 1 kHz tone at 114 dB. The calibrator should be pushed carefully (firmly, but slowly) on to the microphone.

The meter should now be adjusted accordingly. There will often be a free-field adjustment to make for the microphone in use. So, for example, you may actually be adjusting to 113.8 dB on the meter. Meters like the doseBadge Pro make these adjustments automatically and deal with any corrections needed.

Acoustic Calibrator - Class 2, 114 dB

Specifications

Standards	IEC 60942:2003 Class 2 ANSI S1.40:2006	Battery type	1 x 9v (PP3, 6LR61, 6F22)
Sound Level	114.0 dB \pm 0.4 dB	Dimensions	121 x Ø47 mm
Frequency	1 kHz \pm 1%	Weight	220 g with battery
Distortion	< 2%	Controls	Push button start, Automatic switch off with green LED indication. User override of automatic switch off
Reference conditions	23°C, 101.3 kPa, 65% RH	CE classification	EN 61000-6-3:2007+A1:2011 EN 61000-6-1:2007
Microphone size	½" (to IEC 1094-4)		

Head Office

NoiseMeters Inc
3233 Coolidge Hwy
Berkley
MI 48072
USA

Telephone **888 206 4377**
Fax **888 584 2230**

Email: **info@noisemeters.ca**
Support: **support@noisemeters.ca**

Web Sites

Main site:
<https://www.noisemeters.ca>

Product shortcut:
<https://www.noisemeters.ca/product/cr/calibrator/cr518/>

Tech Support:
<https://support.noisemeters.com>